



**AMETEK**<sup>®</sup>  
MATERIALS ANALYSIS DIVISION

## Certificate of Calibration

Document Number MCD-025-A0

Amptek Inc. hereby certifies that the above product was calibrated in accordance with applicable procedures, and meets all published specifications. The unit was calibrated using standards whose accuracy is traceable to the National Institute of Standards and Technology (NIST).

Calibration Date: 25-NOV-2015  
Calibration Due: 25-NOV-2016  
Model Number: MCA8000D  
Description: Multichannel Analyzer  
Serial Number: 791  
Manufacturer: Amptek Inc.  
Calibration Procedure: MCD-510-A0  
Certified By: *Leslie K. Marchbanks*

### Tolerances

V Scale	Offset @ 40 mV	Gain
0-1 V Scale	±0.2 mV	±0.2% @ 0.8 V
0-10 V Scale	±2.0 mV	±0.5% @ 8 V

Received  
Returned

in tol \_\_\_\_\_  
in tol ✓

out tol \_\_\_\_\_  
out tol \_\_\_\_\_

N/A ✓  
N/A \_\_\_\_\_