

CALIBRATION CERTIFICATE**33210****Applicant**

Customer name CCSTEC GesmbH
Contact Angelika Zweimüller
Address Triesterstrasse 36
A-2512 Oeynhausen
Austria

Order reference applicant -
Order reference TPF Control 33209

Instrument information [DUT]

Instrument type Flow Calibrator
Manufacturer Mesa Laboratories
Serial no. 115673
Model Definer 220-H
Tag no. -

Calibration method

The device under test is connected in a parallel setup to the mentioned flow calibrator to compare flow readings. An appropriate warm up time is incorporated.
A flow source is connected to the inlet of the instrument to generate a flow.

Period of calibration 11 January 2016

Calibration result

The results of the calibration are presented on the following page(s). The reported uncertainty of measurement is based on the standard uncertainty multiplied by a coverage factor $k=2$ which provides a coverage probability of approximately 95%. The standard uncertainty of measurement has been determined in accordance with EA-4/02.

Calibration traceability

The measurements have been executed using standards for which the traceability to (inter)national standards has been demonstrated towards the RvA and apply to ISO 17025.

Remarks

Certificate issue date Boven Leeuwen, 11 January 2016

Calibration technician



Bart Vissers

Technical Manager



Rik van de Bovenkamp

TPF Control b.v.
Van Heemstraweg 19
6657KD BOVEN LEEUWEN (NL)
Tel. +31 85 7500 110
Fax. +31 85 7500 111
Web: www.tpf-control.nl

RvA is member of the European Co-operation for Accreditation (EA) and is one of the signatories to the EA Multilateral Agreement and to the ILAC Mutual Recognition Arrangements (MRA) for the mutual recognition of calibration certificates.

Reproduction of the complete certificate is allowed. Parts of the certificate may only be reproduced after written approval of the calibration laboratory. This certificate is issued under the provision that the Raad voor Accreditatie does not assume any liability.

CALIBRATION CERTIFICATE

33210

Calibration conditions

Calibration gas : Nitrogen (grade 5.0)

Note:

1. Temperature and Pressure calibration are not performed under ISO 17025:2005
2. The reported flow values are standardized flows to 101325 Pa and 21.1 °C.

Calibration results

Lab temperature: 21.1 °C

Atmospheric pressure: 985.5 mBar

INSTRUMENT		LAB STANDARD		DEVIATION (ERROR)			UNCERTAINTY	
[mls/min]		[mls/min]	Tool no.	[%] O.R.	[mls/min]	Limit [%]	CMC [%]	Flow [%]
301.55		300.89	B00801	0.22	0.66	1.00	0.18	0.22
5007.1		5015.5	B00803	-0.17	-8.40	1.00	0.18	0.24
29950		29922	B00803	0.09	28.00	1.00	0.18	0.21

INSTRUMENT		LAB STANDARD		DEVIATION (ERROR)			UNCERTAINTY	
[°C]		[°C]	Tool no.	[%] O.R.	[°C]	Limit [°C]	CMC [°C]	Temp [°C]
21.5		21.4	T00101	0.47	0.1	0.8	0.12	0.15

INSTRUMENT		LAB STANDARD		DEVIATION (ERROR)			UNCERTAINTY	
[mmHg]		[mmHg]	Tool no.	[%] O.R.	[mmHg]	Limit [mmHg]	CMC [mmHg]	Pres [mmHg]
738		739	P00107	-0.14	-1.0	3.5	0.68	0.68

The deviation is determined by: $\text{Deviation} = \frac{\text{Instrument reading} - \text{Lab standard reading}}{\text{Lab standard reading}} * 100 \%$