

Analyt-MTC Service-Center

Klosterrunsstrasse 18, D79379 Muellheim
Tel.: +49(0)7631/5545 Fax: +49(0)7631/ 1 47 40
E-Mail: info@analyt-mtc.de

Flow Calibration Certificate

Rep-N° / Customer	12252 / CCS-TEC GMBH
Serial-N°	330530-8
Model	RotMM p03/1
Range (FS)	60 L/min
Accuracy	1.5%FS
Gas	Air
Pressure	1,012.0 mbar
Temperature	21.3 °C
Procedure	FlowCal Ver4.0.0.9 07.09.2016 08:38:34

Calibration Data

MM	Calibrator Reading (L/min)
20	6.5
30	10.6
40	14.5
50	18.5
60	22.3
70	26.5
80	30.6
90	34.7
100	38.8
110	43.1
120	47.4
130	51.6
140	56.0
150	60.0

Calibrators

Calibrator	Serial-N°	Uncertainty	Date due	Cert-N°
CALTOOL#782	83072	0.4%RD+0.2%FS	15.09.2017	14079

All Instruments used in the calibration are certified against traceable standards.

Report-Nr: 14088	Calibration Date: 16.09.2016
Technician: F. Honore	Approved by: H. Roberg

CC Rev.:2.4 7.08.15