



## Calibration certificate

Certificate number: LH01551

Revision 2.1

### Applicant

: CCSTEC GesmbH  
Triesterstrasse 36  
A-2512, Oeynhausen  
Austria

### Instrument

: Description : Airborne particle counter  
: Manufacturer : Lighthouse  
: Model : Remote 2014  
: Serial no. : 110421002  
: Customer ID : C79

### Date of calibration

: 01 March 2016

### Due date calibration

: March 2017

### Calibration location

: Lighthouse Benelux Calibration laboratory  
in Boven-Leeuwen

### Calibration method

: Calibration has been accomplished as described in ISO21501-4:2007.  
All work performed is in accordance with Lighthouse Worldwide  
Solutions Benelux, Master Calibration Document: I201 and is recorded  
and maintained as such.

### Results

: The results of the measurements are shown in the appendix.  
No readjustments are made.  
The unit under test meets its specifications at all the points measured.

### Traceability

: The measurements have been executed using standards for which the  
traceability to (inter)national standards has been demonstrated.

## CALIBRATION SERVICES

A.A.J. van den Heuvel  
Quality & Customer service Coordinator