



Calibration certificate

Certificate number: LH11045

Revision 2.9

Applicant

: CCSTEC GesmbH
Triesterstrasse 36
A-2512, Oeynhausen
Austria

Instrument

: Description : Airborne particle counter
: Manufacturer : Lighthouse
: Model : Remote 2014
: Serial no. : 110421002
: Customer ID : C79

Date of calibration

: 13 February 2017

Due date calibration

: February 2018

Calibration location

: Lighthouse Benelux Calibration laboratory
in Boven-Leeuwen

Calibration method

: Calibration has been accomplished as described in ISO21501-4:2007.
All work performed is in accordance with Lighthouse Worldwide
Solutions Benelux, Master Calibration Document: I201 and is recorded
and maintained as such.

Results

: The results of the measurements are shown in the appendix.
The unit is readjusted.
All measured values comply with the specification limit(s).

Traceability

: The measurements have been executed using standards for which the
traceability to (inter)national standards has been demonstrated towards
the Raad voor Accreditatie.

Uncertainty

: The reported uncertainty is based on a standard uncertainty multiplied
by a coverage factor of $k = 2$, which provides a confidence level of
approximately 95%. The standard uncertainty is determined in
accordance with EA-4/02. The measurement uncertainty is taken into
account for compliance statements.

M.A.W. van Boxtel
Head of calibration



Appendix

Certificate number: LH11045

Date of calibration : 13 February 2017
Calibrated by : M.A.W. van Boxtel
Function : Head of calibration
Environmental conditions : The ambient temperature was 23,1°C
at a relative humidity of 29,8%

Measurement results, As found

Performance information			
Test	Test result	Expanded uncertainty	Criteria
ISO21501-4 Flow	2,85l/min	0,10l/min	2,83l/min $\pm 5\%$
JIS B 9921 Zero count	0	N/A	<1 in 5 mins
ISO21501-4 False count rate	Upper confidence level 212 particles/m3	N/A	N/A
ISO21501-4 Timer check	60,1 seconds	0,40s	$\pm 1\%$
ISO21501-4 Size resolution	5,62%	3,07%SR	< 15%
ISO21501-4 Counting efficiency 50%	36,6% used size 0,20 μ m	2,75%CE	(50 ± 20)%
ISO21501-4 Counting efficiency 100%	97,0% used size 0,30 μ m	6,10%CE	(100 ± 10)%

Size calibration					
Channel	Channel Size	Threshold voltage	Calculated size	Expanded uncertainty	Calculated Sizing error ($\pm 10\%$)
1	0,20 μ m	66mV	0,205 μ m	0,005 μ m	2,47%
2	0,30 μ m	482mV	0,297 μ m	0,006 μ m	-1,10%
3	0,50 μ m	2139mV	0,518 μ m	0,008 μ m	3,64%
4	1,00 μ m	4913mV	1,065 μ m	0,018 μ m	6,48%

Notes

Reference equipment

Model	S/N	Due to	Certificate number
TSI Mass Flowmeter 4043	40431138006	15-Sep-2017	36750
MCA8000D	567	21-Apr-2017	1603-18232
Fluke-175	32820179	1-Jul-2017	1691396
Fisher Scientific 0666256	150783685	20-Oct-2017	1042-7138479
Solair 1100+	060805006	1-May-2017	LH08302

Particle size standards

Size	Description	Due to	Lot#
0,20 μ m	3200 & 3200A Nanosphere size standards	1-Apr-2017	43115
0,30 μ m	3300A & 3300B, Nanosphere Size Standards	1-Feb-2019	164983
0,50 μ m	3500 & 3500A, Nanosphere Size Standards	1-Apr-2019	168223
1,00 μ m	4009A, 4009A & 4009B Microsphere Size Standards	1-Jan-2019	164509



C7P

Appendix

Certificate number: LH11045

Date of calibration : 13 February 2017
Calibrated by : M.A.W. van Boxtel
Function : Head of calibration
Environmental conditions : The ambient temperature was 23,1°C at a relative humidity of 29,8%

Measurement results, As left

Performance information			
Test	Test result	Expanded uncertainty	Criteria
ISO21501-4 Flow	2,85l/min	±0,10l/min	2,83l/min ±5%
JIS B 9921 Zero count	0	N/A	<1 in 5 mins
ISO21501-4 False count rate	Upper confidence level 212 particles/m3	N/A	N/A
ISO21501-4 Timer check	60,0 seconds	0,40s	±1%
ISO21501-4 Size resolution	2,14%	3,37%SR	< 15%
ISO21501-4 Counting efficiency 50%	39,1% used size 0,20µm	3,23%CE	(50±20)%
ISO21501-4 Counting efficiency 100%	101,4% used size 0,30µm	6,28%CE	(100±10)%

Size calibration					
Channel	Channel Size	Threshold voltage	Calculated size	Expanded uncertainty	Calculated Sizing error (±10%)
1	0,20µm	58mV	0,201µm	0,005µm	0,50%
2	0,30µm	519mV	0,303µm	0,006µm	1,00%
3	0,50µm	2075mV	0,508µm	0,008µm	1,60%
4	1,00µm	4552mV	0,994µm	0,018µm	-0,60%

Notes

Reference equipment

Model	S/N	Due to	Certificate number
TSI Mass Flowmeter 4043	40431138006	15-Sep-2017	36750
MCA8000D	567	21-Apr-2017	1603-18232
Fluke-175	32820179	1-Jul-2017	1691396
Fisher Scientific 0666256	150783685	20-Oct-2017	1042-7138479
Solair 1100+	060805006	1-May-2017	LH08302

Particle size standards

Size	Description	Due to	Lot#
0,20µm	3200 & 3200A Nanosphere size standards	1-Apr-2017	43115
0,30µm	3300A & 3300B, Nanosphere Size Standards	1-Feb-2019	164983
0,50µm	3500 & 3500A, Nanosphere Size Standards	1-Apr-2019	168223
1,00µm	4009A, 4009A & 4009B Microsphere Size Standards	1-Jan-2019	164509