


C83



# Certificate of Calibration

Document Number MCD-025-A1

Amptek Inc. hereby certifies that the above product was calibrated in accordance with applicable procedures, and meets all published specifications. The unit was calibrated using standards whose accuracy is traceable to the National Institute of Standards and Technology (NIST).

Calibration Date: 19-JUN-2017

Calibration Due: 19-JUN-2018

Model Number: MCA8000D, Option PA

Description: Multichannel Analyzer

Serial Number: 738

Manufacturer: Amptek Inc.

Calibration Procedure: MCD-510-A0

Certified By: *Linda R. Vandenkerk*

### Tolerances

V Scale	Offset @ 100 mV	Gain
0-1 V Scale	±0.2 mV	±0.2% @ 0.9 V
0-10 V Scale	±2.0 mV	±0.5% @ 9 V

Received in tol ✓ out tol \_\_\_\_\_ N/A \_\_\_\_\_

Returned in tol ✓ out tol \_\_\_\_\_ N/A \_\_\_\_\_