

## Calibration Certificate

44377

### Applicant

Customer name CCSTEC GesmbH  
Address Triesterstrasse 36  
2512 Oeynhausen  
Austria

Order reference applicant CP0999  
Order reference TPF Control 44376

### Instrument information

Manufacturer TSI  
Instrument type Mass Flow Device  
Model 4043 H  
Serial number 4043 1653 004  
Tag Number C87

### Calibration method

The device under test is directly, in line, connected to a flow calibrator to compare flow readings. An appropriate warm up time is incorporated.  
A flow source is connected to the inlet of the instrument to generate a flow.

### Environmental conditions

The laboratory environment was maintained at  $21^{\circ}\text{C} \pm 2^{\circ}\text{C}$  and  $40\%\text{rh} \pm 20\%\text{rh}$ .  
The atmospheric pressure at the time of calibration was 1005 mBar.

Date (or period) of calibration 15 January 2018

### Results

The results of the calibration are presented on the following page(s).

### Uncertainty

The reported expanded uncertainty is based on the standard uncertainty of the measurement multiplied by a coverage factor  $k$ , such that the coverage probability corresponds to approximately 95%.  
The standard uncertainty of measurement has been determined in accordance with EA-4/02.

### Traceability

The measurements have been executed using standards for which the traceability to (inter)national standards has been demonstrated towards the RvA.

Date 15 January 2018

Calibration Technician



Jonny Crum

Technical Manager



Rik van de Bovenkamp

## Calibration Certificate

44377

### Instrument specification [ Device Under Test ]

Process gas : Air  
Qmax : 200 ls/min  
Reference conditions : 21.1 °C & 1013.25 mBar

Input signal : n.a  
Output signal : Display  
Inlet pressure : 0 Bar (g)  
Outlet pressure : 0 Bar (g)


### Calibration conditions

Calibration gas : Air  
Gas conversion factor : 1.0000 ( TSI )

Calibration inlet pressure : 0 Bar (g)  
Calibration outlet pressure : 0 Bar (g)

### Calibration results

Instrument reading			Reference	Deviation (ERROR)			Uncertainty
Output Display	Full scale	Calculated flow	Flow	Of rate	DUT - REF	Limit	Calibration Flow
[ ls/min ]	[%]	[ ls/min ]	[ ls/min ]	[%]	[ ls/min ]	[%]	[ % ]
0.00							
9.98	4.99	9.98	9.9316	0.47	0.05	3.00	0.33
20.00	10.00	20.00	19.765	1.17	0.23	3.00	0.33
50.05	25.03	50.05	50.518	-0.93	-0.47	3.00	0.31
100.26	50.13	100.26	101.04	-0.77	-0.78	3.00	0.32
150.12	75.06	150.12	150.97	-0.56	-0.85	3.00	0.31
200.06	100.03	200.06	200.89	-0.41	-0.83	3.00	0.32



### Note

- The deviation is determined by : 
$$\text{Deviation} = \frac{\text{Instrument reading} - \text{Reference}}{\text{Reference}} \times 100 \%$$
- The hysteresis of the DUT can be determined by the deviation between two series but is not included in the uncertainty.
- The Indicated Lab Standard Flow is derived applying the gas conversion factor. Calibration takes place with the calibration gas, results are reported in process gas (DUT).