

C75

Analyt-MTC Service-Center

Klosterrunsstrasse 18, D79379 Muelheim
Tel.: +49(0)7631/5545 Fax. +49(0)7631/ 1 47 40
E-Mail. info@analyt-mtc.de

Rotameter Calibration Certificate

Calibration Parameters

Rep-N° / Customer	13822/CCS-TEC GMBH
Calibration gas	Air
Process gas	
Standard pressure	1013 hPa
Standard temperature	21,11 °C
Comment	

DUT

Manufacturer	ANALYT-MTC GmbH
Model	RotMM p03/1
SN	330530-8
Flow Range	0 to 60 L/min

Reference

Calibrator	SN	Uncertainty	Date due	Cert-N°
DPC47	496155-1	0.5% RD+0.2%FS	11.02.2020	190211496155-1_000

Calibration Data

Calibrator Rd.	DUT Rd.
(slm@21,11C)	(mm)
5,95	20
10,02	30
14,20	40
18,00	50
22,04	60
26,06	70
30,03	80
34,20	90
38,40	100
42,80	110
47,00	120
51,20	130
55,60	140
59,60	150

Every instrument used in the calibration are certified against traceable standards

Report-Nr:-20190221_002	Calibration date: Feb 21 2019
Technician: Christophe J.	Approved by: H. Roberg

