

C 85



CALIBRATION | SOLUTIONS
TO | IMPROVE | YOUR | PERFORMANCE



CERTIFICATE OF CALIBRATION

1812-12817

Customer information

Client : Cleanroom Control And Sterilization Technology
Contact : Mevr. Angelika Zweimüller
Address : Triesterstrasse 36
2512 Oeynhausen
Germany
Reference client :
Reference Trescal : 201825737/1

Instrument information

Make / type : AMPTEK INC / MCA8000D
Description : Recorder/datalogger
Range :
Serial number : 00791
Identification number : 300001
Accuracy :

Date of calibration : 02 January 2019

Method of calibration

P1-02-E.025 Calibration of multimeters.

Environmental conditions (limits during measurements)

Ambient temperature : $23\text{ °C} \pm 1\text{ °C}$
Relative humidity : 40%rh - 55%rh

Used reference

The equipment used is traceable to National and/or International standards.
150654/1 High performance Calibrator Cert.180612186

Note

The instrument is measured but not adjusted, so the results are both 'as found' as 'as left'.
Recommended recalibration date : 02 January 2020

Conclusion

All reported results are within the accuracy specified by the manufacturer, without taking into account the proven uncertainty.

Issue date: 02 January 2019

Technician
Eelco Ooms

This document may not be reproduced other than in full, except with the prior written approval of the head of the issuing laboratory.
Unless otherwise stated, the calibration was performed at the address mentioned in the footnote.

Trescal bv | Storkstraat 2-4 | 2722 NN Zoetermeer | The Netherlands | T + 31 79 343 00 00 | E info@trescal.nl



* c 1 8 1 2 1 2 8 1 7 *

CERTIFICATE OF CALIBRATION

1812-12817

Conclusion : The results comply with the manufacturer specifications.

Calibration Method : After acclimatising for a period not less than 1 hour; the measurement functions of the MCA800D are verified using a calibrator.

The instrument is not adjusted.

The measured values are the result of 1 observations.

The error is defined as value of the UUT (reading/setting) - value of the Standard(supplied/measured).

Gain accuracy at 1 kHz

UUT Range	Supplied	UUT Reading	Limit Min	Limit Max	Error ± Uncertainty
1 V	800,0 mV	0,79912 V	0,79508 V	0,80499 V	(-0,9 ± 0,4) mV
10 V	8,000 V	8,0009 V	7,9508 V	8,0499 V	(1 ± 4) mV