


Calibration certificate

Certificate number: LH39413

Revision 3.0

Applicant	: CCSTEC GesmbH Triesterstrasse 36 A-2512, Oeynhausen Austria
Instrument	: Description : Airborne particle counter Manufacturer : Lighthouse Model : Solair 3100E Serial no. : 190504038 Customer ID : C96
Date of calibration	: 03 June 2019
Due date calibration	: June 2020
Calibration location	: Lighthouse Benelux Calibration laboratory in Boven-Leeuwen
Calibration method	: Calibration has been accomplished as described in ISO21501-4:2018. All work performed is in accordance with Lighthouse Worldwide Solutions Benelux, Master Calibration Document: I201 and is recorded and maintained as such.
Results	: The results of the measurements are shown in the appendix. The unit is readjusted.
Compliance, As left	: For some of the measured values it is not possible to make a statement of compliance with specification.
Traceability	: The measurements have been executed using standards for which the traceability to (inter)national standards has been demonstrated towards the Raad voor Accreditatie.
Uncertainty	: The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor of $k = 2$, which provides a confidence level of approximately 95%. The standard uncertainty is determined in accordance with EA-4/02. The measurement uncertainty is taken into account for compliance statements according to ILAC-G8:03/2009.

M.A.W. van Boxtel
Head of calibration



Appendix

Certificate number: LH39413

Date of calibration : 03 June 2019
Calibrated by : M.A.W. van Boxtel
Function : Head of calibration
Environmental conditions : The ambient temperature was 23,7°C at a relative humidity of 40,6%

Measurement results. As found

Performance information			
Test	Test result	Expanded uncertainty	Criteria
ISO21501-4 Flow	28,73l/min	0,83l/min	28,30l/min $\pm 5\%$
JIS B 9921 Zero count	0	N/A	≤ 1 in 5 mins
ISO21501-4 False count rate	Upper confidence level 21 particles/m ³	N/A	N/A
ISO21501-4 Timer check	60,0 seconds	0,40s	$\pm 1\%$
ISO21501-4 Size resolution	1,43% used size 0,40 μ m	3,09%SR	$< 15\%$
ISO21501-4 Counting efficiency 50%	57,9% used size 0,30 μ m	2,67%CE	(50 ± 20)%
ISO21501-4 Counting efficiency 100%	97,2% used size 0,50 μ m	10,83%CE	(100 ± 10)%

Size calibration					
Channel	Channel Size	Threshold voltage	Calculated size	Expanded uncertainty	Calculated Sizing error ($\pm 10\%$)
1	0,30 μ m	49mV	0,302 μ m	0,006 μ m	0,72%
2	0,50 μ m	453mV	0,492 μ m	0,008 μ m	-1,50%
3	1,00 μ m	1058mV	0,979 μ m	0,018 μ m	-2,09%
4	3,00 μ m	3223mV	2,767 μ m	0,053 μ m	-7,78%
5	5,00 μ m	3758mV	5,082 μ m	0,167 μ m	1,64%
6	10,00 μ m	4606mV	10,339 μ m	0,222 μ m	3,39%

Notes

Counting efficiency: used size 1,0 μ m: 87,9 %CE $\pm 7,06\%$ CE

Reference equipment

Model	S/N	Due to	Certificate number
TSI Mass Flowmeter 4043	40431350007	22-Mar-2020	52822
MCA8000D	566	26-Apr-2020	1904-13246
Fluke-175	29250596	22-Nov-2019	1811-09376
Fisher Scientific 0666256	170471257	26-Jun-2019	1042-8643931
Palas	6752 / 6751	20-Nov-2019	235-10808

Particle size standards

Size	Description	Due to	Lot#
0,30 μ m	3300A & 3300B, Nanosphere Size Standards	1-Apr-2021	196947
0,40 μ m	3400A, Nanosphere Size Standards	1-Apr-2021	197091
0,50 μ m	3500 & 3500A, Nanosphere Size Standards	1-Jan-2021	193188
1,00 μ m	4009A, 4009A & 4009B Microsphere Size Standards	1-Aug-2021	200992
3,00 μ m	4D-03 Dry Duke Microsphere Size standards	1-Jul-2021	199811
5,00 μ m	4D-05 Dry Duke Microsphere Size standards	1-Aug-2020	188015
10,00 μ m	4D-10 Dry Duke Microsphere Size standards	1-Apr-2021	196944

Appendix

Certificate number: LH39413

Date of calibration : 03 June 2019
 Calibrated by : M.A.W. van Bortel
 Function : Head of calibration
 Environmental conditions : The ambient temperature was 23,7°C at a relative humidity of 40,6%

Measurement results, As left

Performance information				
Test	Test result	Expanded uncertainty	Criteria	
ISO21501-4 Flow	28,73l/min	±0,83l/min	28,30l/min ±5%	
JIS B 9921 Zero count	0	N/A	≤1 in 5 mins	
ISO21501-4 False count rate	Upper confidence level 21 particles/m ³	N/A	N/A	
ISO21501-4 Timer check	60,0 seconds	0,40s	±1%	
ISO21501-4 Size resolution	1,43% used size 0,40µm	3,09%SR	< 15%	
ISO21501-4 Counting efficiency 50%	58,0% used size 0,30µm	2,67%CE	(50±20)%	
ISO21501-4 Counting efficiency 100%	96,4% used size 0,50µm	10,73%CE	(100±10)%	

Size calibration					
Channel	Channel Size	Threshold voltage	Calculated size	Expanded uncertainty	Calculated Sizing error (±10%)
1	0,30µm	48mV	0,301µm	0,006µm	0,43%
2	0,50µm	473mV	0,501µm	0,008µm	0,15%
3	1,00µm	1082mV	1,000µm	0,018µm	-0,04%
4	3,00µm	3330mV	3,004µm	0,078µm	0,12%
5	5,00µm	3743mV	4,999µm	0,164µm	-0,03%
6	10,00µm	4551mV	9,999µm	0,220µm	-0,01%

Notes

Counting efficiency: used size 1,0 µm: 90,3 %CE ± 7,25%CE

Reference equipment

Model	S/N	Due to	Certificate number
TSI Mass Flowmeter 4043	40431350007	22-Mar-2020	52822
MCA8000D	566	26-Apr-2020	1904-13246
Fluke-175	29250596	22-Nov-2019	1811-09376
Fisher Scientific 0666256	170471257	26-Jun-2019	1042-8643931
Palas	6752 / 6751	20-Nov-2019	235-10808

Particle size standards

Size	Description	Due to	Lot#
0,30µm	3300A & 3300B, Nanosphere Size Standards	1-Apr-2021	196947
0,40µm	3400A, Nanosphere Size Standards	1-Apr-2021	197091
0,50µm	3500 & 3500A, Nanosphere Size Standards	1-Jan-2021	193188
1,00µm	4009A, 4009A & 4009B Microsphere Size Standards	1-Aug-2021	200992
3,00µm	4203 and 4203A, Particle Size Standards	1-Sep-2021	202646
5,00µm	4D-05 Dry Duke Microsphere Size standards	1-Aug-2020	188015
10,00µm	4D-10 Dry Duke Microsphere Size standards	1-Apr-2021	196944