

Calibration Certificate**54400****Applicant**

Customer name CCSTEC GesmbH
Address Triesterstrasse 36
2512 Oeynhausen
Austria

Order reference applicant CP0999
Order reference TPF Control 54353

Instrument information

Instrument type Flow Calibrator
Manufacturer Mesalabs
Model Definer 220H
Serial number 115673
Tag number C63

Calibration method

The temperature calibration is done by comparing the DUT reading to a PT-100 reading in Air. The pressure DUT is directly, connected to a pressure standard to compare pressure readings. The flow measurements are made in a parallel setup.

A flow source is connected to the inlet of the instrument to generate a flow.

Environmental conditions

The laboratory environment was maintained at $21^{\circ}\text{C} \pm 2^{\circ}\text{C}$ and $40\%\text{rh} \pm 20\%\text{rh}$.
The atmospheric pressure at the time of calibration was 1007 mBar.

Date (or period) of calibration 5 June 2019 - 6 June 2019

Results The results of the calibration are presented on the following page(s).

Uncertainty The reported uncertainty of measurement is based on the standard uncertainty multiplied by a coverage factor $k=2$, which for a normal distribution corresponds to a coverage probability of approximately 95%.

The standard uncertainty of measurement has been determined in accordance with EA-4/02.

Traceability

The measurements have been executed using standards for which the traceability to (inter)national standards has been demonstrated towards the RvA.

Date 6 June 2019

Calibration Technician



Bart Vissers

Technical Manager



Rik van de Bovenkamp

TPF Control b.v.
Van Heemstraweg 19
6657KD BOVEN LEEUWEN (NL)
Tel. +31 85 7500 110
Fax. +31 85 7500 111
Web: www.tpf-control.nl

RvA is member of the European Co-operation for Accreditation (EA) and is one of the signatories to the EA Multilateral Agreement and to the ILAC Mutual Recognition Arrangements (MRA) for the mutual recognition of calibration certificates.

Reproduction of the complete certificate is allowed.
Parts of the certificate may only be reproduced after written approval of the calibration laboratory.

This certificate is issued under the provision that the Raad voor Accreditatie does not assume any liability.

Calibration Certificate

54400

Instrument specification [Device Under Test]

Qmax : 30000 sccm
Reference conditions : 21.1 °C & 1013.25 mBar

Serial number : 115673
Tag number : C63

Calibration conditions

Calibration gas : Air

Calibration results

Instrument reading	Reference	Deviation (ERROR)			Uncertainty
		Of rate [%]	DUT- REF [sccm]	Limit [%]	Calibration [%]
Flow [sccm]	Flow [sccm]				
301.85	300.43	0.47	1.42	1.00	0.19
5018.6	5004.5	0.28	14.10	1.00	0.19
29953	29866	0.29	87.00	1.00	0.20
Temperature [°C]	Temperature [°C]	Of rate [%]	DUT - REF [°C]	Limit [°C]	Calibration [°C]
21.9	21.7	0.92	0.2	0.8	0.4
Pressure [mBar (a)]	Pressure [mBar (a)]	Of rate [%]	DUT- REF [mBar (a)]	Limit [mBar (a)]	Calibration [mBar (a)]
1008	1008	0.00	0	5	3

Calibration results after adjustment

Instrument reading	Reference	Deviation (ERROR)			Uncertainty
		Of rate [%]	DUT- REF [sccm]	Limit [%]	Calibration [%]
Flow [sccm]	Flow [sccm]				
302.77	302.31	0.15	0.46	1.00	0.19
5009.7	5005.9	0.07	3.80	1.00	0.18
29956	29867	0.30	89.00	1.00	0.18
Temperature [°C]	Temperature [°C]	Of rate [%]	DUT - REF [°C]	Limit [°C]	Calibration [°C]
21.2	21.2	0.00	0.0	0.8	0.4
Pressure [mBar (a)]	Pressure [mBar (a)]	Of rate [%]	DUT- REF [mBar (a)]	Limit [mBar (a)]	Calibration [mBar (a)]
1007	1007	0.00	0	5	3

Note

- The deviation is determined by :
$$\text{Deviation} = \frac{\text{Instrument reading} - \text{Reference}}{\text{Reference}} \cdot 100 \%$$
- Calibrations are performed at mentioned pressure and temperature conditions. Reference temperatures are defined according the ITS-90.