

CALIBRATION CERTIFICATE

2005-06547

Customer information

Client : Cleanroom Control And Sterilization Technology
Contact : Mevr. Angelika Zweimüller
Address : Triesterstrasse 36
2512 Oeynhausen
Germany
Reference client : CP0999790152
Reference Trescal : 202019660/1

Instrument information

Make / type : AMPTEK INC / MCA8000D
Description : Recorder/datalogger
Range :
Serial number : 00738
Identification number : C83
Accuracy :

Date of calibration : 15 May 2020

Method of calibration

P1-02-E.001 Calibration of electrical measurement instruments

Electrical instruments are calibrated with a multifunctional calibrator and/or a digital multimeter.

Environmental conditions (used in determination of uncertainty)

Ambient temperature : 23 °C ± 1 °C
Relative humidity : 40%rh - 55%rh

Used reference

The measurements have been executed using standards for which the traceability to (inter)national standards has been demonstrated towards the Dutch Accreditation Council (RvA).
150654/2 High performance Calibrator Cert.181116156

Note

The instrument is measured but not adjusted, so the results are both 'as found' as 'as left'.

Technician
Eelco Ooms

Issue date: 15 May 2020

Head of the laboratory
Luc Van Pelt



RvA is member of the European Co-operation for Accreditation (EA) and is one of the signatories of the EA Multilateral Agreement and to the ILAC (International Laboratory Accreditation Co-operation) Mutual Recognition Arrangements (MRA) for the mutual recognition of calibration certificates.
This certificate is issued provided that the Dutch Accreditation Council does not assume any liability. This document is issued in accordance with the conditions for accreditation of the RvA which is based on ISO/IEC 17025. This document may not be reproduced other than in full, except with the prior written approval of the head of the issuing laboratory.
Unless otherwise stated, the calibration was performed at the address mentioned in the footnote.

Trescal Zoetermeer bv | Storkstraat 2-4 | 2722 NN Zoetermeer | The Netherlands | E info@trescal.nl



CALIBRATION CERTIFICATE

2005-06547

Conclusion : The results comply with the manufacturer specifications.

Calibration Method : After acclimatising for a period not less than 1 hour; the measurement functions of the MCA800D are verified using a calibrator.

The instrument is not adjusted.

The measured values are the result of 1 observation.

The error is defined as value of the UUT (reading/setting) - value of the Standard(supplied/measured).

Gain accuracy at 1 kHz

UUT Range	Supplied	UUT Reading	Limit Min	Limit Max	Error ± Uncertainty
1 V	800,0 mV	0,80290 V	0,79508 V	0,80499 V	(2,9 ± 0,4) mV
10 V	8,000 V	8,0011 V	7,9508 V	8,0499 V	(1 ± 4) mV

The stated uncertainty is that of the entire set-up including the object under test.

The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a level of confidence of approximately 95% .

The uncertainty is calculated following EA-4/02 in accordance with the requirements of the ISO/IEC 17025.