



LIGHTHOUSE
WORLDWIDE SOLUTIONS BENELUX BV



Calibration certificate

Certificate number: LH62134

Revision 3.3

Applicant

: CCSTEC GesmbH
Triesterstrasse 36
A-2512, Oeynhausen
Austria

Instrument

: Description : Airborne particle counter
: Manufacturer : Lighthouse
: Model : Remote 2014
: Serial no. : 191121036

Customer ID : C97

Date of calibration

: 14 December 2020

Due date calibration

: December 2021

Calibration location

: Lighthouse Benelux Calibration laboratory
in Boven-Leeuwen

Calibration method

: Calibration has been accomplished as described in ISO21501-4:2018. All work performed is in accordance with Lighthouse Worldwide Solutions Benelux, Master Calibration Document: I201 and is recorded and maintained as such.

Results

: The results of the measurements are shown in the appendix.
The unit is readjusted.

Compliance, As left

: All measured values comply with the specification limit(s).

Traceability

: The measurements have been executed using standards for which the traceability to (inter)national standards has been demonstrated towards the Raad voor Accreditatie.

Uncertainty

: The reported expanded uncertainty is based on the standard uncertainty of the measurement multiplied by a coverage factor k, such that the coverage probability corresponds to approximately 95%. The standard uncertainty is determined in accordance with EA-4/02. Compliance statements are given in accordance with ISO21501-4:2018.

Date of issue

: 15 December 2020

Authorized by

: M.A.W. van Boxel
Head of calibration

Lighthouse Worldwide Solutions Benelux BV
Van Heemstraweg 19A 6657 KD
Boven-Leeuwen The Netherlands
Tel: +31 (0)487 560811
E-mail: csbenelux@golighthouse.com

The Raad voor Accreditatie is one of the signatories of the Multilateral Agreement of the European Cooperation for Accreditation (EA) for the mutual recognition of calibration certificates.

Reproduction of the complete certificate is allowed, parts of the certificate may only be reproduced with written approval of the calibration laboratory. This certificate is issued provided that neither Lighthouse Worldwide Solutions Benelux BV nor the Raad voor Accreditatie assumes any liability.



Appendix

Certificate number: LH62134

Date of calibration : 14 December 2020

Calibrated by : A.C.G. Gerritse

Function : Calibration & Service Engineer

Environmental conditions : The ambient temperature was 21,8°C
at a relative humidity of 40,9%

Measurement results, As found

Performance information

Test	Test result	Expanded uncertainty	Criteria	
ISO21501-4 Flow	2,95l/min	0,11l/min	2,83l/min $\pm 5\%$	Pass
JIS B 9921 Zero count	0	N/A	≤ 1 in 5 mins	Pass
ISO21501-4 False count rate	Upper confidence level 212 particles/m3	N/A	N/A	N/A
ISO21501-4 Timer check	60,1 seconds	0,40s	$\pm 1\%$	Pass
ISO21501-4 Size resolution	1,11% used size 0,24 μ m	5,27%SR	$< 15\%$	Pass
ISO21501-4 Counting efficiency 50%	49,1% used size 0,20 μ m	2,59%CE	(50 ± 20)%	Pass
ISO21501-4 Counting efficiency 100%	100,2% used size 0,30 μ m	5,65%CE	(100 ± 10)%	Pass

Size calibration

Channel	Channel Size	Threshold voltage	Calculated size	Expanded uncertainty	Calculated Sizing error ($\pm 10\%$)	
1	0,20 μ m	49mV	0,199 μ m	0,007 μ m	-0,66%	Pass
2	0,30 μ m	484mV	0,306 μ m	0,007 μ m	2,16%	Pass
3	0,50 μ m	2251mV	0,523 μ m	0,010 μ m	4,53%	Pass
4	1,00 μ m	4022mV	1,031 μ m	0,018 μ m	3,09%	Pass

Notes

Counting efficiency: used size 0,5 μ m: 99,1 %CE $\pm 5,68\%$ CE
Counting efficiency: used size 1,0 μ m: 101,3 %CE $\pm 6,05\%$ CE

Reference equipment

Model	S/N	Due to	Certificate number
TSI Mass Flowmeter 4040	40401919024	12-Aug-2021	63031
MCA8000D	564	3-Mar-2021	20-13808-3-1
Fluke-175	12170245	6-Apr-2021	20-13808-4-2
Fisher Scientific 0666256	200359247	16-Jun-2022	1042-11351217
Palas	6752 / 6751	23-Nov-2021	235-11058

Particle size standards

Size	Description	Due to	Lot#
0,20 μ m	3200 & 3200A Nanosphere size standards	1-Mar-2022	D20014
0,25 μ m	3240A, Nanosphere size standards	1-Apr-2021	D20024
0,30 μ m	3300A & 3300B, Nanosphere Size Standards	1-Apr-2023	D20015
0,50 μ m	3500 & 3500A, Nanosphere Size Standards	1-Apr-2022	D20017
1,00 μ m	4009A, 4009A & 4009B Microsphere Size Standards	1-Oct-2022	D20020



Appendix

Certificate number: LH62134

Date of calibration : 14 December 2020

Calibrated by : A.C.G. Gerritse
Function : Calibration & Service Engineer

Environmental conditions : The ambient temperature was 21,8°C
at a relative humidity of 40,9%

Measurement results. As left

Performance information

Test Light	Test result	Expanded uncertainty	Criteria	
ISO21501-4 Flow	2,95l/min	±0,11l/min	2,83l/min ±5%	Pass
JIS B 9921 Zero count	0	N/A	≤1 in 5 mins	Pass
ISO21501-4 False count rate	Upper confidence level 212 particles/m3	N/A	N/A	N/A
ISO21501-4 Timer check	60,1 seconds	0,40s	±1%	Pass
ISO21501-4 Size resolution	0,00% used size 0,24µm	5,79%SR	< 15%	Pass
ISO21501-4 Counting efficiency 50%	39,8% used size 0,20µm	2,10%CE	(50±20)%	Pass
ISO21501-4 Counting efficiency 100%	96,6% used size 0,30µm	5,45%CE	(100±10)%	Pass

Size calibration

Channel	Channel Size	Threshold voltage	Calculated size	Expanded uncertainty	Calculated Sizing error (±10%)	
1	0,20µm	52mV	0,204µm	0,007µm	1,75%	Pass
2	0,30µm	454mV	0,300µm	0,007µm	0,09%	Pass
3	0,50µm	2162mV	0,501µm	0,010µm	0,27%	Pass
4	1,00µm	4033mV	1,000µm	0,018µm	0,01%	Pass

Notes

Counting efficiency: used size 0,5 µm: 96,5 %CE ± 5,53%CE
Counting efficiency: used size 1,0 µm: 94,5 %CE ± 5,64%CE

Reference equipment

Model	S/N	Due to	Certificate number
TSI Mass Flowmeter 4040	40401919024	12-Aug-2021	63031
MCA8000D	564	3-Mar-2021	20-13808-3-1
Fluke-175	12170245	6-Apr-2021	20-13808-4-2
Fisher Scientific 0666256	200359247	16-Jun-2022	1042-11351217
Palas	6752 / 6751	23-Nov-2021	235-11058

Particle size standards

Size	Description	Due to	Lot#
0,20µm	3200 & 3200A Nanosphere size standards	1-Mar-2022	D20014
0,25µm	3240A, Nanosphere size standards	1-Apr-2021	D20024
0,30µm	3300A & 3300B, Nanosphere Size Standards	1-Apr-2023	D20015
0,50µm	3500 & 3500A, Nanosphere Size Standards	1-Apr-2022	D20017
1,00µm	4009A, 4009A & 4009B Microsphere Size Standards	1-Oct-2022	D20020