

Calibration Certificate

66078

Applicant

Customer name CCSTEC GesmbH
Address Triesterstrasse 36
2512 Oeynhausen
Austria

Order reference applicant CP0999
Order reference TPF Control 66021

Instrument information

Manufacturer TSI
Instrument type Volume Flow Device
Model 4043 H
Serial number 4043 1653 004
Tag number C87

Calibration method

The device under test is directly, in line, connected to a flow calibrator to compare flow readings. An appropriate warm up time is incorporated.
A flow source is connected to the inlet of the instrument to generate a flow.

Environmental conditions

The laboratory environment was maintained at $21^{\circ}\text{C} \pm 2^{\circ}\text{C}$ and $40\%\text{rh} \pm 20\%\text{rh}$.
The atmospheric pressure at the time of calibration was 1016 mBar.

Date (or period) of calibration 8 January 2021

Results

The results of the calibration are presented on the following page(s).

Uncertainty

The reported uncertainty of measurement is based on the standard uncertainty multiplied by a coverage factor $k=2$, which for a normal distribution corresponds to a coverage probability of approximately 95%.

The standard uncertainty of measurement has been determined in accordance with EA-4/02.

Traceability

The measurements have been executed using standards for which the traceability to (inter)national standards has been demonstrated towards the RvA.

Date 8 January 2021

Calibration Technician



Jonny Crum

Technical Manager



Rik van de Bovenkamp

TPF Control b.v.
Van Heemstraweg 19
6657KD BOVEN LEEUWEN (NL)
Tel. +31 85 7500 110
Fax. +31 85 7500 111
Web: www.tpf-control.nl

RvA is member of the European Co-operation for Accreditation (EA) and is one of the signatories to the EA Multilateral Agreement and to the ILAC Mutual Recognition Arrangements (MRA) for the mutual recognition of calibration certificates.

Reproduction of the complete certificate is allowed.
Parts of the certificate may only be reproduced after written approval of the calibration laboratory.
This certificate is issued under the provision that the Raad voor Accreditatie does not assume any liability.

Calibration Certificate

66078

Instrument specification [Device Under Test]

Process gas : Air
Qmax : 200 l/min

Input signal : n.a
Output signal : Display
Inlet pressure : 0 Bar (g)
Outlet pressure : 0 Bar (g)


Calibration conditions

Calibration gas : Air
Gas conversion factor : 1.0000 (TSI)

Calibration inlet pressure : 0 Bar (g)
Calibration outlet pressure : 0 Bar (g)

Calibration results

Instrument reading			Reference	Deviation (ERROR)			Uncertainty
Output Display	Full scale	Calculated flow	Flow	Of rate	DUT - REF	Limit	Calibration Flow
[l/min]	[%]	[l/min]	[l/min]	[%]	[l/min]	[%]	[%]
0.00							
10.04	5.02	10.04	9.9200	1.25	0.12	2.00	0.52
20.01	10.00	20.01	19.823	0.93	0.19	2.00	0.34
50.06	25.03	50.06	50.117	-0.12	-0.06	2.00	0.31
100.0	49.99	100.0	100.42	-0.44	-0.42	2.00	0.33
149.9	74.96	149.9	151.11	-0.79	-1.21	2.00	0.31
199.9	99.97	199.9	202.12	-1.08	-2.22	2.00	0.30



Note

- The deviation is determined by :
$$\text{Deviation} = \frac{\text{Instrument reading} - \text{Reference}}{\text{Reference}} \times 100 \%$$
- The hysteresis of the DUT can be determined by the deviation between two series but is not included in the uncertainty.
- The indicated Lab Standard Flow is derived applying the gas conversion factor. Calibration takes place with the calibration gas, results are reported in process gas (DUT).
- Calibration done with inlet filter.