

Calibration complies with ISO/IEC 17025, ANSI/NCSL Z540-1, and 9001



Cert. No.: 1051-13063161

Traceable® Certificate of Calibration for Jumbo-Digit Stopwatch

Manufactured for and distributed by: VWR International LLC Researchpark Haasrode 2020, Geldenaaksebaan 464, B - 3001 Leuven

Instrument Identification:

Model: 609-0211,

S/N: 221211461

Manufacturer: Control Company

Standards/Equipment:

Description

Serial Number

Due Date

NIST Traceable Reference

Non-Contact Frequency Counter

26,66879

29 Jul 2022

1000470098

Certificate Information:

Technician: 447

Procedure: CAL-01

Cal Date: 02 Feb 2022

Cal Due Date: 02 Feb 2024

Test Conditions:

51.81%RH 24.48°C 1011mBar

Calibration Data: (New Instrument)

Unit(s)	Nominal	As Found	In Tol	Nominal	As Left	In Tol	Min	Max	±U	TUR
sec/24hr	N.A.	N.A.		0.000	0.133	Y	-0.864	0.864	0.041	>4:1

This certificate indicates Traceability to standards provided by (NIST) National Institute of Standards and Technology and/or a National Standards Laboratory.

A Test Uncertainty Ratio of at least 4:1 is maintained unless otherwise stated and is calculated using the expanded measurement uncertainty. Uncertainty evaluation includes the instrument under test and is calculated in accordance with the ISO "Guide to the Expression of Uncertainty in Measurement: (GUM). The uncertainty represents an expanded uncertainty using a coverage factor k=2 to approximate a 95% confidence level. In tolerance conditions are based on test results falling within specified limits with no reduction by the uncertainty of the measurement. The results contained herein relate only to the item calibrated. This certificate shall not be reproduced except in full, without written approval of Control Company.

Nominal=Standard's Reading; As Left=Instrument's Reading; In Tol=In Tolerance; Min/Max=Acceptance Range; ± U=Expanded Measurement Uncertainty; TUR=Test Uncertainty Ratio; Accuracy=±(Max-Min)/2; Min=As Left Nominal(Rounded) – Tolerance; Max= As Left Nominal(Rounded) + Tolerance;

Rich Rodriguez

Nicol Rodriguez, Quality Manager

Marisa Elms, Technical Manager

lote :

Maintaining Accuracy:

In our opinion once calibrated your Jumbo-Digit Stopwatch should maintain its accuracy. There is no exact way to determine how long calibration will be maintained. Jumbo-Digit Stopwatch change little, if any at all, but can be affected by aging, temperature, shock, and contamination.

Recalibration:

For factory calibration and re-certification traceable to National Institute of Standards and Technology contact Control Company.

Issue Date: 02 Feb 2022

CONTROL COMPANY 12554 Galveston RD Suite B230 Webster TX USA 77598
Phone 281 482-1714 Fax 281 482-9448 sales@control3.com www.traceable.com

Control Company is an ISO/IEC 17025:2005 Calibration Laboratory Accredited by (A2LA) American Association for Laboratory Accreditation, Certificate No. 1750.01.

Control Company is ISO 9001:2015 Quality Certified by DNV GL, Certificate No. CERT-01805-2006-AQ-HOU-ANAB.

International Laboratory Accreditation Cooperation - Multilateral Recognition Arrangement (ILAC-MRA).