

## NANOSPHERE™ SIZE STANDARDS NIST Traceable Mean Diameter

**1. DESCRIPTION.** These particle size standards provide accurate and traceable size calibration for particle size analysis. They are part of a series of polymer microspheres with calibrated mean diameters traceable to the Standard Meter through the National Institute of Standards and Technology (NIST). Diameters from 20 nanometers (nm) to 160 micrometers (µm) are available as aqueous suspensions in dropper-tipped vials, calibrated by photon correlation spectroscopy (PCS), transmission electron microscopy (TEM) or optical microscopy. The aqueous medium has been prepared to promote dispersion and reduce clumping of the particles. The approximate particle concentration in percent solids is given to facilitate dilution for the calibration and validation of particle analyzers. Diameters from 200 µm to 1000 µm are available as dry spheres, calibrated by optical microscopy. The certified mean diameter is traceable to NIST. Other values are for information only and should not be used as calibration values.

**2. PHYSICAL DATA.**

Certified Mean Diameter:	Catalog Number: 3700 and 3700A, Nominal 700 nm
Standard Deviation:	702 nm ± 6 nm, k=2
Coefficient of Variation:	4.9 nm
Microsphere Composition:	0.7%
Microsphere Density:	Polystyrene
Index of Refraction:	1.05 g/cm <sup>3</sup>
Approximate Concentration:	1.59 @ 589 nm
	1% solids

- Continued on page 2

### CERTIFICATE OF CALIBRATION AND TRACEABILITY

This certifies that the calibrated mean diameter was transferred by transmission electron microscopy (TEM) from the National Institute of Standards and Technology (NIST) certified microspheres (Standard Reference Material 1963, 1691 or 1690).

Catalog Number: 3700 and 3700A, Nanosphere™ Size Standards

Certification Date:	September 16, 2015
Certified Batch:	3700-004
Production Batch:	3700-026
Certified Mean Diameter:	702 nm
Expanded Uncertainty:	± 6 nm, k=2



*Joe Vasiliou* 9.23.15  
Joe Vasiliou, Senior Metrology Engineer  
Thermo Fisher Scientific Particle Technology

Packaging Lot # 163612

Expiration Date: JAN'19